




Polycare Examination Gloves,
Powder-free Documentation

Meditech

Content

1. Polycare Product Sheet with Picture
2. CE Marking
3. 510(k)
4. FDA Certificate
5. EN 420, EN 16523 & EN 374-4
6. EN 455 1 & 3
7. ASTM D 6499
8. ISO 13485: 2016 Certificate
9. Declaration of Conformity
10. Manufacturing License
11. Carton on a Pallet Bird View
12. Image of Polycare Gloves
13. Image of Box
14. Image of Carton

1. Polycare Product Sheet with Picture



2. CE Marking



By Royal Charter

EU Type Examination Certificate

This is to certify that:

Meditech Gloves Sdn. Bhd.
PT 3345, Jalan Permata 1/3
Arab Malaysian Industrial Park
Nilai
Negeri Sembilan, Darul Khusus
71800
Malaysia

Holds Certificate Number:

CE 698568

In respect of:

**Natural rubber gloves for personal protection- models MEPF3 & MSPF2
(NBR) Nitrile Butadiene Rubber Latex gloves for personal protection- model MNEPF1.
To EN 420:2003+A1:2009, EN ISO 374-1:2016 & EN ISO 374-5:2016.**

on the basis that BSI carried out the relevant Type Examination procedures under the requirements with the Regulation (EU) 2016/425 of the European Parliament and Council relating to Personal Protective Equipment Regulation (PPE) Annex V (Module B) and meets the relevant health and safety requirements specified in Annex II

For and on behalf of BSI, a Notified
Body for the above Regulation
(Notified Body Number 0086):

Chris Lewis - Certification Director, Product Certification

First Issued: 2019-05-02
Latest Issue: 2019-07-31

Effective Date: 2019-07-31
Expiry Date: 2024-05-02

Page: 1 of 6

...making excellence a habit.™

This certificate has been issued by and remains the property of BSI Assurance UK Ltd, Kitemark Court, Davy Avenue, Knowlhill, Milton Keynes MK5 8PP, United Kingdom and should be returned immediately upon request.
To check its validity telephone +44 (0) 345 080 9000. An electronic certificate can be authenticated [online](#).

BSI Assurance UK Limited, registered in England under number 7805321 at 389 Chiswick High Road, London W4 4AL, UK.
A member of BSI Group of Companies.

EU Type Examination Certificate

No. CE 698568

Product Specification:

The PPE equipment Protective Gloves that are covered by the scope of this Module B Certificate and the Harmonized European Standards and technical specifications to which the products are approved are to the following specifications:

Model:	MEPF3 Polycare Latex Examination Gloves
Classification:	Protective gloves for use against microorganisms and general applications.
Description:	Natural rubber latex examination gloves, powder free and non-sterile. Ambidextrous gloves with beaded cuff, available in off-white to light yellow colours.
Size Range:	XS to XL
Product codes:	MEPF3-XS, MEPF3-S, MEPF3-M, MEPF3-L and MEPF3-XL.

Product Specification

Performance: **Resistance to penetration to EN 374-2:2014**
Pass

Resistance to degradation to EN 374-4:2013
Tested for degradation against the chemical listed below.
Sodium Hydroxide 40% (CAS:1310-73-2)
Mean Degradation: -84.3%

Resistance to chemical permeation to EN ISO 374-1:2016 (Test method EN 16523-1:2015) **Type C**

Chemical	Level
Sodium Hydroxide 40% (K)	6

General requirements for gloves to EN 420:2003+A1:2009

Dexterity: Level 5
pH: 7.5
Protein Content: 84.5 µg/g

First Issued: 2019-05-02
Latest Issue: 2019-07-31

Effective Date: 2019-07-31
Expiry Date: 2024-05-02

Page: 2 of 6

This certificate has been issued by and remains the property of BSI Assurance UK Ltd, Kitemark Court, Davy Avenue, Knowlhill, Milton Keynes MK5 8PP, United Kingdom and should be returned immediately upon request.
To check its validity telephone +44 (0) 345 080 9000. An electronic certificate can be authenticated [online](#).

BSI Assurance UK Limited, registered in England under number 7805321 at 389 Chiswick High Road, London W4 4AL, UK.
A member of BSI Group of Companies.

EU Type Examination Certificate

No. CE 698568

Product Specification continued:

Model:	MSPF2 - ExcellPC Sterile Latex Surgical Gloves Powder Free
Classification:	Protective gloves for use against microorganisms and general applications.
Description:	Natural rubber latex sterile polymer coated Surgical Gloves, powder free, disposable single use, sterilized by Gamma Irradiation, beaded cuff and anatomical shape., available in off-white to light yellow colours.
Size Range:	5.5 to 9.0
Product codes:	MSPF2-5.5, MSPF2-6.0, MSPF2-6.5, MSPF2-7.0, MSPF2-7.5, MSPF2-8.0, MSPF2-8.5 and MSPF2-9.0.

Product Specification

Performance: **Resistance to penetration to EN 374-2:2014**
Pass

Resistance to degradation to EN 374-4:2013
Tested for degradation against the chemical listed below.
Sodium Hydroxide 40% (CAS:1310-73-2)
Mean Degradation: -78.9%

Resistance to chemical permeation to EN ISO 374-1:2016 (Test method EN 16523-1:2015) **Type C**

Chemical	Level
Sodium Hydroxide 40% (K)	6

General requirements for gloves to EN 420:2003+A1:2009

Dexterity: Level 5
pH: 7.2
Protein Content: 123.5 µg/g

First Issued: 2019-05-02

Latest Issue: 2019-07-31

Effective Date: 2019-07-31

Expiry Date: 2024-05-02

Page: 3 of 6

This certificate has been issued by and remains the property of BSI Assurance UK Ltd, Kitemark Court, Davy Avenue, Knowlhill, Milton Keynes MK5 8PP, United Kingdom and should be returned immediately upon request.
To check its validity telephone +44 (0) 345 080 9000. An electronic certificate can be authenticated [online](#).

BSI Assurance UK Limited, registered in England under number 7805321 at 389 Chiswick High Road, London W4 4AL, UK.
A member of BSI Group of Companies.

EU Type Examination Certificate

No. CE 698568

Product Specification continued:

Model:	MNEPF1 – BPG Nitrile Examination Gloves Powder-free
Classification:	Protective gloves for use against microorganisms and general applications.
Description:	NBR (Nitrile Butadiene Rubber Latex) examination gloves, powder free and non-sterile. Ambidextrous gloves with beaded cuff, available in blue colour.
Size Range:	XS to XL
Product codes:	MNEPF1-XS MNEPF1-S, MNEPF1-M, MNEPF1-L and MNEPF1-XL.

Product Specification

Performance: **Resistance to penetration to EN 374-2:2014**
Pass

Resistance to degradation to EN 374-4:2013
Tested for degradation against the chemical listed below.
Sodium Hydroxide 40% (CAS:1310-73-2)
Mean Degradation: -21.6%

Resistance to chemical permeation to EN ISO 374-1:2016 (Test method EN 16523-1:2015) **Type C**

Chemical	Level
Sodium Hydroxide 40% (K)	6

General requirements for gloves to EN 420:2003+A1:2009

Dexterity: Level 5
pH: 6.2
Protein Content: N/A

First Issued: 2019-05-02

Latest Issue: 2019-07-31

Effective Date: 2019-07-31

Expiry Date: 2024-05-02

Page: 4 of 6

This certificate has been issued by and remains the property of BSI Assurance UK Ltd, Kitemark Court, Davy Avenue, Knowlhill, Milton Keynes MK5 8PP, United Kingdom and should be returned immediately upon request.
To check its validity telephone +44 (0) 345 080 9000. An electronic certificate can be authenticated [online](#).

BSI Assurance UK Limited, registered in England under number 7805321 at 389 Chiswick High Road, London W4 4AL, UK.
A member of BSI Group of Companies.

EU Type Examination Certificate

No. CE 698568

Applicable Standards:

The following Harmonized European Standards:

EN 420:2003+A1:2009 Protective gloves. General requirements.

EN ISO 374-1:2016. Protective gloves against dangerous chemicals and micro-organisms. Terminology and performance requirements for chemical risks.

EN 374-2:2014. Protective gloves against dangerous chemicals and microorganisms. Determination of resistance to penetration.

EN 374-4:2013. Protective gloves against chemicals and micro-organisms. Determination of resistance to degradation by chemicals.

EN ISO 374-5:2016. Protective gloves against dangerous chemicals and micro-organisms. Terminology and performance requirements for micro-organisms risks.

EN 16523-1:2015. Determination of material resistance to permeation by chemicals. Permeation by liquid chemical under conditions of continuous contact.

First Issued: 2019-05-02

Latest Issue: 2019-07-31

Effective Date: 2019-07-31

Expiry Date: 2024-05-02

Page: 5 of 6

This certificate has been issued by and remains the property of BSI Assurance UK Ltd, Kitemark Court, Davy Avenue, Knowlhill, Milton Keynes MK5 8PP, United Kingdom and should be returned immediately upon request.
To check its validity telephone +44 (0) 345 080 9000. An electronic certificate can be authenticated [online](#).

BSI Assurance UK Limited, registered in England under number 7805321 at 389 Chiswick High Road, London W4 4AL, UK.
A member of BSI Group of Companies.

EU Type Examination Certificate

No. CE 698568

Certificate Administration Details

Manufacturer's technical file reference: MEPF3 v1, MSPF2 v0 and MNEPF1 v0 Technical Files.

Certificate Amendment Record:

Issue Date	Comments	Internal BSI Project Number
April 2019	Transition of PPE Directive, Article 10 CE 684183. Model: MEPF3. Addition on new models MNEPF1 & MSPF2.	0086:19:9640615
July 2019	Addition of sizes XS & XL to MNEPF1	0086:19:3043355

Note: The Certificate holder is responsible for ensuring that the Notified Body is advised of changes to any aspect of the overall processes utilised in the manufacture of the product, failure to do so could invalidate the Certificate in respect of product manufactured following the introduction of such changes.

Monitoring of manufactured PPE:

The validity of the Certificate is also dependent on the conformity to the type based on the internal production control plus supervised product checks at random intervals (Annex VII, Module C2), for the specific standards/product that are referenced in the BSI issued Module C2 Certificate Number CE 615886.

First Issued: 2019-05-02

Latest Issue: 2019-07-31

Effective Date: 2019-07-31

Expiry Date: 2024-05-02

Page: 6 of 6

This certificate has been issued by and remains the property of BSI Assurance UK Ltd, Kitemark Court, Davy Avenue, Knowlhill, Milton Keynes MK5 8PP, United Kingdom and should be returned immediately upon request.
To check its validity telephone +44 (0) 345 080 9000. An electronic certificate can be authenticated [online](#).

BSI Assurance UK Limited, registered in England under number 7805321 at 389 Chiswick High Road, London W4 4AL, UK.
A member of BSI Group of Companies.



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
 10903 New Hampshire Avenue
 Document Control Room -WO66-G609
 Silver Spring, MD 20993-0002

Dr. Effendi Tenang
 Managing Director
 Best Putra Gloves SDN BHD
 Lot 1357-1358, JLN Kg. Mohd Taib
 Kawasan Perindustrian Sg. Choh
 48000 Rawang, Selangor Darul Ehsan, Malaysia

NOV 10 2010

Re: K101105

Trade/Device Name: Powder-Free, Latex Examination Glove, Natural Color,
 Non-Sterile, Model: MEPF1
 Regulation Number: 21 CFR 880.6250
 Regulation Name: Patient Examination Glove
 Regulatory Class: I
 Product Code: LYY
 Dated: October 4, 2010
 Received: October 21, 2010

Dear Dr. Tenang:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Page 2- Dr. Tenang

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourceforYou/Industry/default.htm>.

Sincerely yours,



fe

Anthony D. Watson, B.S., M.S., M.B.A.
Director
Division of Anesthesiology, General Hospital,
Infection Control and Dental Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure



BEST PUTRA GLOVES SDN.BHD. (580515-T)

Lot 1357 & 1358, Jln Kg. Mohd Taib, Kawasan Perindustrian 5g, Choh
49000 Rawang, Selangor Darul Ehsan, Malaysia.
Tel: +03-6092 1042, +03-6092 1142 Fax: +03-6091 2820
E-mail: bestputra@bpg.com Website: www.bpggloves.com



K101105

NGV 10

INDICATIONS FOR USE

510(k) Number: K101105

Device Name: Powder – Free, Latex Examination Glove, Natural Color, Non-Sterile, Model: MEPF1

Indications For Use:

A powder-free patient examination glove is a disposable device intended for medical purposes that is worn on the examiner's hand or finger to prevent contamination between patient and examiner.

Signature:

Dr. Effendi Tepak
(Managing Director)

Prescription Use _____
(Part 21 CFR 801 Subpart D)

OR

Over – The – Counter Use XXX
(21 CFR 801 Subpart C)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Page 1 of ____

(Division Sign-off)
Division of Anesthesiology, General Hospital
Infection Control, Dental Devices

510(k) Number: K101105

4. FDA Certificate

The screenshot shows the FDA website's 'Establishment Registration & Device Listing' page. The header includes the U.S. Department of Health & Human Services logo and the FDA logo. A search bar is located in the top right corner. The main navigation menu includes links for Home, Food, Drugs, Medical Devices, Radiation-Emitting Products, Vaccines, Blood & Biologics, Animal & Veterinary, Cosmetics, and Tobacco Products. The page title is 'Establishment Registration & Device Listing'. Below the title, there are breadcrumb links for 'FDA Home', 'Medical Devices', and 'Databases'. A 'New Search' button and a 'Back To Search Results' link are also present. The main content area displays the following information:

Proprietary Name:	BPG Latex Examination Gloves - Powder-Free; POLYCARE Powder Free Latex Examination Gloves , MEPF3
Classification Name:	LATEX PATIENT EXAMINATION GLOVE
Product Code:	LYY
Device Class:	1
Regulation Number:	880.6250
Medical Specialty:	General Hospital
Registered Establishment Name:	MEDITECH GLOVES SDN BHD
Registered Establishment Number:	3008484177
Premarket Submission Number:	K101105
Owner/Operator:	MEDITECH GLOVES SDN BHD
Owner/Operator Number:	10033273
Establishment Operations:	Manufacturer

Page Last Updated: 09/21/2020
Note: If you need help accessing information in different file formats, see [Instructions for Downloading Viewers and Players](#).
Language Assistance Available: Español | 繁體中文 | Tiếng Việt | 한국어 | Tagalog | Русский | العربية | Kreyòl Ayisyen | Français | Polski | Português | Italiano | Deutsch | 日本語 | فارسی | English

The footer contains the FDA logo and links for Accessibility, Contact FDA, Careers, FDA Basics, FOIA, No FEAR Act, Nondiscrimination, and Website Policies.

[FDA Website-Link](#)



SATRA Technology Centre Ltd
Wyndham Way, Telford Way, Kettering,
Northamptonshire, NN16 8SD United Kingdom
Tel: +44 (0) 1536 410000
Fax +44 (0) 1536 410626
email: info@satra.co.uk
www.satra.co.uk



Customer details: BSI Group
PO Box 6221
Kitemark Court
Davy Avenue
Milton Keynes
MK1 9EP

SATRA reference: SPC0253226/1701/
SMcD/A

Your reference: 4500095154.

Date of report: 1st March 2017

Samples received: 5th January 2017

For the attention of: Kinga Demetriou

Date(s) work
carried out: 16th to 22nd February
2017

TECHNICAL REPORT

Subject: Chemical innocuousness testing in accordance with BS EN 420: 2003 + A1:
2009 on gloves described as MEPF3

Conditions of Issue:

This report may be forwarded to other parties provided that it is not changed in any way. It must not be published, for example by including it in advertisements, without the prior, written permission of SATRA.


Results given in this report refer only to the samples submitted for analysis and tested by SATRA. Comments are for guidance only.


Tests marked # fall outside the UKAS Accreditation Schedule for SATRA. All interpretations of results of such tests and the comments based upon them are outside the scope of UKAS accreditation and are based on current SATRA knowledge.

A satisfactory test report in no way implies that the product tested is approved by SATRA and no warranty is given as to the performance of the product tested. SATRA shall not be liable for any subsequent loss or damage incurred by the client as a result of information supplied in the report.


The uncertainty of the results (UoM) in this report is based on a standard uncertainty multiplied by a coverage factor $k=2$, which provides for a confidence level of approximately 95%.

Report signed by: S McDonald
Position: Chemical Technologist
Department: Chemical & Analytical Technology





TERMS AND CONDITIONS OF BUSINESS



Issue Date: 1st October 2009



SATRA Technology Centre Ltd
Wyndham Way, Telford Way, Kettering,
Northamptonshire, NN16 8SD United Kingdom
Tel: +44 (0) 1536 410000
Fax +44 (0) 1536 410626
email: info@satra.co.uk
www.satra.co.uk



Customer details: BSI Group
PO Box 6221
Kitemark Court
Davy Avenue
Milton Keynes
MK1 9EP

SATRA reference: SPC0253226/1701/
SMcD/B

Your reference: 4500095154.

Date of report: 2nd March 2017

Samples received: 5th January 2017

For the attention of: Kinga Demetriou

Date(s) work
carried out: 20th to 22nd February
2017

TECHNICAL REPORT

Subject: BS EN 16523-1:2015 resistance to permeation by chemicals on gloves
described as MEPF3 against 40% sodium hydroxide

Conditions of Issue:

This report may be forwarded to other parties provided that it is not changed in any way. It must not be published, for example by including it in advertisements, without the prior, written permission of SATRA.

Results given in this report refer only to the samples submitted for analysis and tested by SATRA. Comments are for guidance only.


Tests marked \neq fall outside the UKAS Accreditation Schedule for SATRA. All interpretations of results of such tests and the comments based upon them are outside the scope of UKAS accreditation and are based on current SATRA knowledge.


A satisfactory test report in no way implies that the product tested is approved by SATRA and no warranty is given as to the performance of the product tested. SATRA shall not be liable for any subsequent loss or damage incurred by the client as a result of information supplied in the report.

The uncertainty of the results (UoM) in this report is based on a standard uncertainty multiplied by a coverage factor $k=2$, which provides for a confidence level of approximately 95%.

Report signed by: S McDonald
Position: Chemical Technologist
Department: Chemical & Analytical Technology

WORK REQUESTED:






APPENDICES:



Gloves described as MEPF3



TERMS AND CONDITIONS OF BUSINESS



Issue Date: 1st October 2009



SATRA Technology Centre Ltd
Wyndham Way, Telford Way, Kettering,
Northamptonshire, NN16 8SD United Kingdom
Tel: +44 (0) 1536 410000
Fax +44 (0) 1536 410626
email: info@satra.co.uk
www.satra.co.uk



Customer details: BSI Group
PO Box 6221
Kitemark Court
Davy Avenue
Milton Keynes
MK1 9EP

SATRA reference: SPC0253226 /1701 /1

Your reference: 4500095154.

Date of report: 6 March 2017

Samples received: 5 January 2017

For the attention of: Kinga Demetriou

Date(s) work carried out: 2 March 2017

TECHNICAL REPORT

Subject: Testing of gloves identified as MEPF3 in accordance with EN 420: 2003 + A1: 2009 clauses 5.1 Sizing and 5.2 Dexterity only and EN 374-2: 2014.

Conditions of Issue:

This report may be forwarded to other parties provided that it is not changed in any way. It must not be published, for example by including it in advertisements, without the prior, written permission of SATRA.

Results given in this report refer only to the samples submitted for analysis and tested by SATRA. Comments are for guidance only.



Tests marked \neq fall outside the UKAS Accreditation Schedule for SATRA. All interpretations of results of such tests and the comments based upon them are outside the scope of UKAS accreditation and are based on current SATRA knowledge.

A satisfactory test report in no way implies that the product tested is approved by SATRA and no warranty is given as to the performance of the product tested. SATRA shall not be liable for any subsequent loss or damage incurred by the client as a result of information supplied in the report.

The uncertainty of the results (UoM) in this report is based on a standard uncertainty multiplied by a coverage factor $k=2$, which provides for a confidence level of approximately 95%.


Report signed by: Adam Mortiboys
Position: PPE Technologist
Department: Safety Product Testing

Work Requested



MEPF3

Conclusion




Note ▲ – Where gloves do not meet the minimum length requirements specified in Table 1 of EN 420:2003 + A1:2009, the standard therefore requires that the manufacturer shall clearly state in the user instructions the intended application of the gloves and the reason why the gloves do not conform to the minimum length requirements.

Note ◆ – As per clause 4.3 of EN 374-2:2014, the gloves submitted for testing were found to be unsuitable for the Air leak test. Therefore as per EN 374-2 only the Water leak test has been performed.



Testing




EN 374-2: 2014 Test Results



TERMS AND CONDITIONS OF BUSINESS

- 1.
- 2.
- 3.
- 4.
- 5.
- 6.
- 7.
- 8.
- 9.
- 10.
- 11.
- 12.
- 13.



Issue Date: 1st October 2009





SURGICAL MATERIAL TESTING LABORATORY

TEST REPORT

Examination Glove Report

Report No: 19/5988/1

Report Date: Wednesday 6th November, 2019

Authors:
Louise Barry
Fara Shadroo

Revision Information:
Template Version: 339a63c
Revision: 1.4
Revision date: Wednesday 6th
November, 2019
Revision Author: louise

The information contained in this report is proprietary, and is not for circulation without the consent of the Surgical Material Testing Laboratory (SMTL), or the commissioning authority/company. This report shall not be reproduced except in full without the written approval of SMTL, Princess of Wales Hospital, Coity Road, Bridgend, CF31 1RQ.

Examination Glove Report

Report No: 19/5988/1

Louise Barry
Fara Shadroo

Wednesday 6th November, 2019

1 Name & Address of Client/Requesting Authority

Paddy Smyth (Best Putra)
Managing Director (Best Putra)
Wan Nadzifah Binti Wan Hassan
QA Manager (Acting)
Meditech Gloves Sdn Bhd
PT 3345, Jln. Permata 1/3,
Arab Malaysian Industrial Park,
71800 Nilai, Negeri Sembilan,
Malaysia

Email: pjsmyth@bpgloves.co.uk
nadzifah@meditechgloves.com.my
michael@bpgloves.co.uk
aminkassim@gmail.com

2 Introduction

This document presents the results of POLYCARE Latex Examination Gloves tested to BS EN 455 Parts 1⁽¹⁾ and 3 Total Extractable Protein⁽²⁾.

3 Test Products/Samples for this project

Table 1: Samples

Supplier	Product Name	Description	Catalogue Number	Batch/Lot Number	Quantity	Date received	SMTL Sample ID
Meditech Gloves	POLYCARE	Latex Examination Gloves Powder Free, Size Medium, DOM 09-2019	MEPF3-M	31938010 9	400	04/10/2019	60869

NOTES:

- The test results in this report relate only to the test sample(s) analysed.
- The Manufacturer, Product Name, Description, Catalogue & Batch Numbers were provided by the client.

3.1 Departures/Abnormalities of Sample Condition

None.

4 Date of Testing


8th - 10th October 2019.

5 Location of Testing


All testing was performed at SMTL premises

6 Testing Details

6.1 Perforations - TM-22⁽³⁾



6.2 Extractable protein - TM-230⁽⁴⁾



6.3 Standards relevant to the test method



6.4 Testing conditions

6.4.1 Perforations

- [Redacted]

6.4.2 Total Extractable Protein

- [Redacted]

6.5 Deviations/exclusions from, and additions to standard methods

6.5.1 Perforations

- None.

The information contained in this report is proprietary, and is not for circulation without the consent of the Surgical Material Testing Laboratory (S. M. T. L.), or the commissioning authority/company. This report shall not be reproduced except in full without the written approval of S. M. T. L., Princess of Wales Hospital, Coity Road, Bridgend, CF31 1RQ.

6.5.2 Total Extractable Protein

- None.

6.6 Uncertainty of Measurement

Uncertainty of measurement (UoM) has not been taken into account when interpreting the test results.

6.7 Sampling Details

All samples were selected and supplied by the client.

The batch size of the gloves supplied was not stated by the client. In accordance with BS EN 455 Part 1, a batch size between 35,001 to 150,000 was chosen, and therefore 50 gloves per stage were tested for perforations using General Inspection Level I at an AQL of 1.5%. With reference to Table 3, the sample size was tested up to the fifth sampling stage or until compliance or non compliance was determined.

6.8 Sample Preparation


Samples were prepared according to the relevant test method used.

7 Results

7.1 Perforation Testing

The results of perforation testing are presented in Table 2. Compliance has been determined with reference to Table 3.


Table 2: Perforation Testing results for Sample 60869

A large black rectangular redaction box covering the content of Table 2.

Note:

- Perforations were detected in the palm (4 gloves), finger (2 gloves), thumb (2 gloves) and 1 glove had 4 perforations in the palm.


Table 3: Multiple sampling - Perforation compliance (BS EN 455-1)

A large black rectangular redaction box covering the content of Table 3.

7.2 Total Extractable Protein

The total amount of extractable protein per gram for each of the inner and outer surface test samples is presented in Table 4.

Table 4: Total Extractable Protein results for Sample 60869



† Ovalbumin Equivalent Protein

8 Authorisation

Approved and signed electronically. Please see last page of this document.

Pete Phillips, Director, SMTL.

References

- (1) *Medical gloves for single use - part 1: Requirements and testing for freedom from holes. BS EN 455-1:2000.*
- (2) *Medical glove for single use - part 3: Requirements and testing for biological evaluation. Annex A: Method for the determination of aqueous extractable proteins in natural rubber gloves using the modified lowry assay. BS EN 455-3:2015.*
- (3) SMTL. *Detection of perforations in medical gloves to BS EN 455 Part 1. (TM-22).*
- (4) SMTL. *Determination of extractable protein in natural rubber gloves. (TM-230).*

The information contained in this report is proprietary, and is not for circulation without the consent of the Surgical Material Testing Laboratory (S. M. T. L.), or the commissioning authority/company. This report shall not be reproduced except in full without the written approval of S. M. T. L., Princess of Wales Hospital, Coity Road, Bridgend, CF31 1RQ.



LEMBAGA GETAH MALAYSIA
MALAYSIAN RUBBER BOARD
RUBBER RESEARCH INSTITUTE OF MALAYSIA
Global Testing and Consultancy for Rubber (G-TACr)

Malaysian Rubber Board, 47000 Sungai Buloh, Selangor.

Tel: (6)03-61459471 Fax: (6)03-61412907

Email: gtacr@lgm.gov.my Website: http://www.lgm.gov.my/gtacr



September 6, 2016

Our Report No : LGM/BTK/UPB/5.10/CP/1609/0004
 Your Reference :

MEDITECH GLOVES SDN BHD
 Lot 3345, Jalan Permata 1/3,
 Arab Malaysia Industrial Park,
 71800 Nilai, Negeri Sembilan Darul Khusus,
 Malaysia.

Fax : 606-7997749

Attention : Nurul Izwana Azman


ANALYTICAL TEST REPORT OF EXTRACTABLE ANTIGENIC CONTENT

We have received samples delivered to us under your cover letter dated July 28, 2016. Analysis had been carried out as per your request. Samples were extracted with Phosphate Buffered Saline (5 ml PBS per gram sample) following the extraction procedure of ASTM D5712-10. We report the following results :

Date of receipt of sample : September 1, 2016
 Date Tested : September 6, 2016
 Test Method : ASTM D6499-07
 Sample Condition : Unaged
 Type of Sample : Glove

Sample Reference	Protein Content µg/g
1. Latex Examination Gloves, Powder-Free (Polymer Coated) Lot no. : 316290110	ND

Note : ND means not detected at limit of detection of 1 µg/g


 Nurul Izwana Azman
 (IKM No.: A2542/5226/07)
 Research Officer
 Malaysian Rubber Board
 For Director General



LEMBAGA GETAH MALAYSIA
MALAYSIAN RUBBER BOARD
RUBBER RESEARCH INSTITUTE OF MALAYSIA
Global Testing and Consultancy for Rubber (G-TACr)

Malaysian Rubber Board, 47000 Sungai Buloh, Selangor.

Tel: (6)03-61459471 Fax: (6)03-61412907

Email: gtacr@lgm.gov.my Website: http://www.lgm.gov.my/gtacr



Our report no. : LGM/BTK/UPB/5.10/CP/1609/0005

Your reference : -

MEDITECH GLOVES SDN. BHD.
PT 3345, JALAN PERMATA 1/3,
71800 NILAI, NEGERI SEMBILAN DARUL KHUSUS, MALAYSIA,
(ATTN : NURUL IZWANA AZMAN)

Fax No.: 06-677 9620



ANALYTICAL RESULT

We have received one sample from you under covering letter dated 28/07/2016 . The sample was analyzed as per your request. The results are reported below.

Sample Description : LATEX EXAMINATION GLOVES,
POWDER FREE (POLYMER COATED)

LOT NO : 316290110

Date Received : 01/09/2016






LEMBAGA GETAH MALAYSIA
MALAYSIAN RUBBER BOARD
RUBBER RESEARCH INSTITUTE OF MALAYSIA
Global Testing and Consultancy for Rubber (G-TACr)
Malaysian Rubber Board, 47000 Sungai Buloh, Selangor.
Tel: (6)03-61459471 Fax: (6)03-61412907
Email: gtacr@lgm.gov.my Website: <http://www.lgm.gov.my/gtacr>



Our report no. : LGM/BTK/UPB/5.10/CP/1609/0005

"BERKHIDMAT UNTUK NEGARA"


Kartini Alias
(IKM NO : M2541/5225/07)
Research Officer
Malaysian Rubber Board
For Director-General

Date of reporting
28 SEPTEMBER 2016

8. ISO 13485: 2016 Certificate



By Royal Charter

Certificate of Registration

QUALITY MANAGEMENT SYSTEM - ISO 13485:2016 & EN ISO 13485:2016

This is to certify that:

Meditech Gloves Sdn. Bhd.
Lot 3345, Jalan Permata 1/3
Arab Malaysian Industrial Park
71800 Nilai
Negeri Sembilan, Darul Khusus
Malaysia

Holds Certificate No:

MD 631470

and operates a Quality Management System which complies with the requirements of ISO 13485:2016 & EN ISO 13485:2016 for the following scope:

The manufacture of powdered and powder-free sterile latex surgical gloves, powder-free sterile latex surgical under gloves and examination gloves.

For and on behalf of BSI:

Stewart Brain, Head of Compliance & Risk - Medical Devices

Original Registration Date: 2015-03-09

Effective Date: 2018-03-09

Latest Revision Date: 2019-02-27

Expiry Date: 2021-03-08

Page: 1 of 1



...making excellence a habit.™

This certificate was issued electronically and remains the property of BSI and is bound by the conditions of contract.

An electronic certificate can be authenticated [online](#).

Printed copies can be validated at www.bsi-global.com/ClientDirectory or telephone +603 2242 4211.

Further clarifications regarding the scope of this certificate and the applicability of ISO 13485:2016 & EN ISO 13485:2016 requirements may be obtained by consulting the organization.

This certificate is valid only if provided original copies are in complete set.

Information and Contact: BSI, Kitemark Court, Davy Avenue, Knowlhill, Milton Keynes MK5 8PP. Tel: + 44 345 080 9000
BSI Assurance UK Limited, registered in England under number 7805321 at 389 Chiswick High Road, London W4 4AL, UK.
A Member of the BSI Group of Companies.

9. Declaration of Conformity

MEDITECH GLOVES SDN BHD (580515-T)
Registration No. 200201012852

PT 3345, JALAN PERMATA 1/3,
ARAB MALAYSIAN INDUSTRIAL PARK,
71800 NILAI,
NEGERI SEMBILAN DARUL KHUSUS,
MALAYSIA
TEL: +6 06 799 7742
+6 06 799 7746
FAX: +6 06 799 7749
http: www.meditechgloves.com

MEDICAL DEVICE DIRECTIVE EC DECLARATION OF CONFORMITY

Manufacturer/ Supplier : **MEDITECH GLOVES SDN. BHD.**
**PT 3345, Jalan Permata 1/3, Arab Malaysian Industrial Park,
71800 Nilai, Negeri Sembilan, Malaysia**

Authorised Representative: **Best Putra Gloves (UK) Ltd.,
103 Carrickasticken Road, Forkhill,
Newry, County Down, BT35 9RL,
Northern Ireland, United Kingdom.**

Model : **Polycare Latex Examination Gloves, Powder-free (MEPF3)**

Classification : **Class I**


Description : **Natural rubber latex examination gloves, powder-free & non-sterile.
Ambidextrous gloves with beaded cuff, available in available in off-white
to light yellow colours.**

Standards Applied:

EN 374-2:2014	EN ISO 374-1:2016	EN 455-1:2020	EN 455-3:2015
EN 374-4:2013	EN 420:2003+A1:2009	EN 455-2:2015	EN 455-4:2009

MEDITECH GLOVES SDN. BHD. declares that the device described above conforms to the relevant provision of **Directive 93/42/EEC**, and complies with the relevant Essential Requirements of the Annex V and is manufactured in accordance with the **ISO 13485: 2016** Medical Device Quality Management System (QMS).

Signed for and on behalf of the manufacturer, Meditech. Gloves. Sdn. Bhd.:


DR. MOHAMMED EFFENDI MOHAMMED TENANG

Managing Director

Meditech Gloves Sdn. Bhd.

18/9/2020

10. Manufacturing License

ASAI,
ORIGINAL

Borang ICA / (Part 1/08)
Form ICA 7 (Rev. 1/08)

KEMENTERIAN
PERDAGANGAN
ANTARABANGSA DAN
INDUSTRI MALAYSIA



MINISTRY OF
INTERNATIONAL TRADE
AND INDUSTRY OF
MALAYSIA

AKTA PENYELARASAN PERINDUSTRIAN, 1975
LESEN PENGILANGAN
INDUSTRIAL CO-ORDINATION ACT, 1975
MANUFACTURING LICENCE

No. Siri: **A 030190**
Serial No.:

No. Lesen: **A 017184**
Licence No.:

PADA MENURUT AKTA PENYELARASAN PERINDUSTRIAN, 1975
IN PURSUANCE OF THE INDUSTRIAL CO-ORDINATION ACT, 1975

Saya memberi lesen kepada.....
I license

MEDITECH GLOVES SDN. BHD.

(Nama pemohon/Applicant's name)

untuk bertindak sebagai PENGILANG BERLESEN mulai dari
to act as LICENSED MANUFACTURER as from
23 April 2009
(Tarikh/Date)

di tempat pengilangan beralamat.....
at the place of manufacturing at

**PT 3345, Jalan Permata 1/3, Arab Malaysian
Industrial Park, 71800 Nilai, Negeri Sembilan Darul Khusus**


bagi keluaran yang dinyatakan di bawah ini dan tertakluk kepada syarat yang dilampirkan bersama ini:
for product(s) specified hereunder and subject to the conditions attached herewith:

Keluaran:
Product:

'Surgical Gloves' dan 'Examination Gloves'


Bertarikh pada
Dated this

/i Ogos 2010



DATUK DR. REBECCA FATIMA STA MARIA
Tindakan Ketua Setiausaha (Perdagangan)
Kementerian Perdagangan Antarabangsa & Industri
Malaysia

11. Carton on a Pallet Bird View


Surgical Carton on pallet bird view




Type A (Current Packing)



Type B (Suggest Packing)



Exam Carton on pallet bird View



12. Image of Polycare Gloves






13. Image of Box





Návod na použitie: Jednorazové vyšetrovacie a ochranné rukavice z prírodného kaučuku (NR), bez obsahu púdru, nesterilné.



NA
Bániky 419/1
IČO:454539

